



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2021-09-29</b>
<b>KIT DESCRIPTION:</b>	<b>BLOOD GAS CONTROL LEVEL 1</b>
<b>CATALOGUE NO:</b>	<b>BG5001</b>
<b>BATCH NO:</b>	<b>248BG</b>
<b>EXPIRY DATE:</b>	<b>2023-09-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>LEVEL I / 248BG</b>	<b>SODIUM</b>	<b>124.00</b>	<b>118.00 -130.00</b>	<b>126.77</b>	<b>mmol/L</b>
<b>LEVEL I / 248BG</b>	<b>POTASSIUM</b>	<b>2.59</b>	<b>2.38 – 2.80</b>	<b>2.71</b>	<b>mmol/L</b>

APPROVED BY:

*H. Martin*

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mar. R J Fitzgerald W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.