

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-03-01
KIT DESCRIPTION:	LIQUID CRP CONTROL (CRP CONTROL)
CATALOGUE NO:	CP2481
BATCH NO:	2948CP
EXPIRY DATE:	2024-08-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 2948CP	CRP	148.4	118.7 – 178.1	151.7	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY: 

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Director: Dr S P Fitzgerald BSc PhD DSc (Hons) FAE Man. R J Fitzgerald W Fitzgerald OBE R F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland Company Number NI15738, VAT Number GB353030400.