

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-01-29
KIT DESCRIPTION:	LIPOPROTEIN(a) CALIBRATOR SET
CATALOGUE NO:	LP3404
BATCH NO:	633015
EXPIRY DATE:	2026-01-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 2968CH	LP(a)	24.4	19.5-29.3	24.4	mg/dL
LEVEL III/ 2970CH	LP(a)	32.6	26.1-39.1	31.5	mg/dL

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MSc R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15739, VAT Number GB 353 0304 00.