

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFAC	CTURE: 2023-04-29		
KIT DESCRIPTION:	IMMUNOASSAY SPECIALITY II – LEVEL 3		
CATALOGUE NO:	IAS3119		
LOT NUMBERS:	2289EC		
EXPIRY DATE:	2025-04-28		
SYSTEM OF ANALY	SIS: ROCHE COBAS e801		

TEST CONTROLS OR VIALS

	RANGE	TARGET	RESULT	PRECISION (if applicable)
2289EC	231 – 385	308	304	N/A
			302	
2289EC	11.1 – 18.5	14.8	15.9	N/A
			15.8	
				+
				2289EC 11.1 – 18.5 14.8 15.9

QC TES Davina (TED BY: Graham		
DATE:	26-Oct-23		
RESULT	(PASS/FAIL):	PASS	
QC APP	ROVED BY:	G. Marfin!	

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Umited is a company registered in Northern Ireland. Company Number NII 5738. VAT Number GB 353 0304 00.



Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R. J Fitzgerald, W Fitzgerald OBE, R. P. Kelly BSc FCA, Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00,