

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

2021-06-29 CALIBRATION SERUM LEVEL 3 (CAL 3)		
1297UE		
2025-06-28		

Analyte ALBUMIN	Lot No. 1297UE	Target 31.10	Range		Result	Units
			0.00	N/A	30.73	g/L
TOTAL PROTEIN	1297UE	47.50	0.00	N/A	47.13	g/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By:

Date

29 Jun 2023.

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom +44 (0) 28 9442 2413 | marketing@randox.com | randox.com

mey