



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2023-01-29</b>
<b>KIT DESCRIPTION:</b>	<b>HIGH SENSITIVITY CRP CONTROL (hsCRP CONTROL)</b>
<b>CATALOGUE NO:</b>	<b>CP2477</b>
<b>BATCH NO:</b>	<b>2946CP</b>
<b>EXPIRY DATE:</b>	<b>2025-01-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>Level II / 2946CP</b>	<b>hsCRP</b>	<b>4.09</b>	<b>3.27-4.91</b>	<b>4.15</b>	<b>mg/l</b>
<b>N/A</b>	<b>N/A</b>	<b>N/A</b>	<b>N/A</b>	<b>N/A</b>	<b>N/A</b>

**APPROVED BY:**  SignNow e-signature ID: 6a9bf4e634...  
15/03/2023 20:16:52 UTC  
Sharon Mumbi (Signer)

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom**  
**T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.