

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2023-05-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS – LEVEL 1
CATALOGUE NO: IA3109
LOT NUMBERS: 2269EC
EXPIRY DATE: 2026-05-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	2269EC	7.87 – 11.8	9.84	9.6	N/A
				9.6	
Total T4 (nmol/l)	2269EC	27.0 – 45.0	36.0	36.5	N/A
				36.6	
Total T3 (ng/dl)	2269EC	62.5 – 104	83.3	86.9	N/A
				84.8	

QC TESTED BY: Kelly Jennings
DATE: 07-Mar-24
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016