

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-12-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 1  
CATALOGUE NO: IA3109  
LOT NUMBERS: 2302EC  
EXPIRY DATE: 2025-12-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA125 (U/ml)	2302EC	14.0 – 21.0	17.5	17	N/A
				17	
CA15-3 (U/ml)	2302EC	18.7 – 28.1	23.4	24	N/A
				25	
CA19-9 (U/ml)	2302EC	18.6 – 28.0	23.3	21.4	N/A
				22.1	

### QC TESTED BY:

Davina Graham

DATE: 06-Jul-23

RESULT (PASS/FAIL): PASS

QC APPROVED BY:

*R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016