

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2023-04-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 2270EC  
EXPIRY DATE: 2026-04-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000

## TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Growth Hormone (uU/ml)	2270EC	14.8 – 22.2	18.5	19.9	N/A
				20.2	
Total T4 (nmol/l)	2270EC	85.5 – 143	114	126	N/A
				126	
T Uptake (%U)	2270EC	26.2 – 37.6	31.9	34.1	N/A
				34.9	

## QC TESTED BY:

Kelly Jennings

DATE: 15-Apr-24

RESULT (PASS/FAIL): PASS

QC APPROVED BY:

*G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016