

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2024-08-29
KIT DESCRIPTION:	LIQUID CRP CONTROL (CRP CONTROL)
CATALOGUE NO:	CP2481
BATCH NO:	3045CP
EXPIRY DATE:	2026-02-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 3045CP	CRP	157	126 - 188	161	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY: 

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MInst R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15739, VAT Number GB 353 0904 00