

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2023-05-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 2271EC  
EXPIRY DATE: 2026-05-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000

## TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Growth Hormone (uU/ml)	2271EC	37.8 – 56.6	47.2	48.9	N/A
				49.2	
Total T4 (nmol/l)	2271EC	143 – 239	191	189	N/A
				199	
T Uptake (%U)	2271EC	31.1 – 44.7	37.9	41.6	N/A
				42.7	

## QC TESTED BY:

Kelly Jennings

DATE: 01-May-24

RESULT (PASS/FAIL): PASS

QC APPROVED BY:

*R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016