

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2023-11-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS – LEVEL 2
CATALOGUE NO: IA3110
LOT NUMBERS: 2342EC
EXPIRY DATE: 2026-11-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Progesterone (nmol/l)	2342EC	26.2 – 39.4	32.8	31.5	N/A
				31.9	
SHBG (nmol/l)	2342EC	28.2 – 42.4	35.3	36.0	N/A
				36.3	
Prolactin (µU/ml)	2342EC	559 – 839	699	714	N/A
				716	

QC TESTED BY: Emma Muir
DATE: 23-Jul-24
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mar. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016