

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-06-29
KIT DESCRIPTION:	HAEMOGLOBIN HbA1c CONTROL SET
CATALOGUE NO:	HA5072
BATCH NO:	686207
EXPIRY DATE:	2026-06-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 2433HA	HbA1c	6.13	4.90-7.36	6.11	%HbA1c
LEVEL II / 2435HA	HbA1c	13.8	11.0-16.6	12.5	%HbA1c

APPROVED BY: You Playa Pulikunni Vijayan.

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com