

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2024-11-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS – LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 2482EC  
EXPIRY DATE: 2027-11-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA125 (U/ml)	2482EC	69.4 – 104	86.8	91.0	N/A
				91.0	
CA15-3 (U/ml)	2482EC	28.2 – 42.2	35.2	36.0	N/A
				36.0	
CA19-9 (U/ml)	2482EC	58.6 – 87.8	73.2	75.0	N/A
				74.5	

QC TESTED BY: Kelly Jennings  
DATE: 14-May-25  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII5738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016