

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2024-11-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS – LEVEL 2
CATALOGUE NO: IA3110
LOT NUMBERS: 2527EC
EXPIRY DATE: 2027-11-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA125 (U/ml)	2527EC	69.4 – 104	86.8	89.0	N/A
				88.0	
CA15-3 (U/ml)	2527EC	28.2 – 42.2	35.2	34.0	N/A
				34.0	
CA19-9 (U/ml)	2527EC	58.6 – 87.8	73.2	72.0	N/A
				67.9	

QC TESTED BY: Davina Graham
DATE: 02-May-25
RESULT (PASS/FAIL): PASS
QC APPROVED BY: A. Mentem

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MSc, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016