



# HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Radox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2025-03-01
KIT DESCRIPTION:	TRI-LEVEL CARDIAC CONTROL (CRD CONTROL 1, 2, 3)
CATALOGUE NO:	CQ3259
BATCH / LOT NO:	760188
EXPIRY DATE:	2029-02-28

ANALYTE	TARGET	RANGE	RESULT	UNITS
4920CK/ CK-NAC	77	62-92	76	U/L
4921CK / CK-NAC	208	166-250	199	U/L
4922CK / CK-NAC	535	428-642	499	U/L
4921CK / CK-MB	17	14-20	15	U/L
4922CK / CK-MB	92	74-110	87	U/L

*Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.*

APPROVED BY: Mol  
MALAIKA COLA CO

**QUALITY CONTROL REPRESENTATIVE**

**Randex Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S. P. Fitzgerald BSc PhD DSc (Hon) FAE Harv. B. J. Fitzgerald, W. Fitzgerald CBE, R.P. Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15734, VAT Number: GB 353 0304 00.

4705-QC Rev6  
27<sup>TH</sup> September 2013