



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-10-29
KIT DESCRIPTION:	Lipid Control (LPD CONTROL 3)
CATALOGUE NO:	LE2670
BATCH NO:	3102CH
EXPIRY DATE:	2026-10-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 3102CH	HDL- Cholesterol	2.06	1.75-2.37	1.89	mmol/l
Level III / 3102CH	LDL- Cholesterol	5.37	4.56-6.18	5.38	mmol/l

APPROVED BY:

*M. Colaco*



Verified by signNow  
20-Jun-2024 16:03:14 UTC  
8056292bba94c0d921e  
Malika Colaco

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738; VAT Number GB353030400