



QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2023-12-29
Product Description:	Liquid Assayed Specific Protein Control (SP Control)
Catalogue Number:	PS2684
Batch Number:	688LPC
Expiry Date:	2025-12-28

Analyte	Lot No.	Target	Range		Result	Units
CRP	688LPC	65.70	52.60	78.80	68.27	mg/l
C4	688LPC	55.00	44.00	66.00	51.57	mg/dl

Approved By:


MALAIKA COLACO

Date: 27-Mar-25

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
+44 (0) 28 9442 2413 | marketing@randox.com | randox.com

Directors: Dr S P FitzGerald CBE FREng DL BSc PhD DSc (Hons) FIAE Man. R P Kelly BSc FCA.
Randox Laboratories Ltd is a subsidiary of Randox Holdings Limited a company registered within Northern Ireland with company number NI. 614690. VAT Registered Number: GB 151 6827 08.

PS2684 Ch
Revision 0

12874-QC
Revision (0)
Last Printed Date: 27/03/2025