



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2025-05-29
Product Description:	ASSAYED URINE CONTROL
Catalogue Number:	AU2353
Batch Number:	1400UC
Expiry Date:	2029-05-28

Analyte	Lot No.	Target	Range		Result	Units
CALCIUM	1400UC	4.86	4.37	5.35	4.79	mmol/l

Approved By: *GBowles*

Date: *22-OCT-25*

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
+44 (0) 28 9442 2413 | marketing@randox.com | randox.com

Directors: Dr S P FitzGerald CBE FREng DL BSc PhD DSc (Hons) FIAE Man. R P Kelly BSc FCA

Randox Laboratories Ltd is a subsidiary of Randox Holdings Limited a company registered within Northern Ireland with company number NI1 614690. VAT Registered Number: GB 151 6827 08.

AU2353 Ch
Revision 0

7677-QC
Revision (3)
Last Printed Date: 22/10/2025