



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-03-29
KIT DESCRIPTION:	Direct LDL/HDL Cholesterol Calibrator
CATALOGUE NO:	CH2673
BATCH NO:	3217CH
EXPIRY DATE:	2026-03-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
3217CH	HDL	1.48	N/A	1.48	mmol/l
3217CH	LDL	3.86	N/A	3.71	mmol/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FHEA FRCR R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI5758, VAT Number GB 353 0304 00.