

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-07-29
KIT DESCRIPTION: Immunoassay Premium Plus Level 3(IA
PREMIUM PLUS 3) China
CATALOGUE NO: IA3111
LOT NUMBERS: 2545EC
EXPIRY DATE: 2028-07-28
SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH mU/mL	2545EC	42.2-63.4	52.8	51.4	N/A
	2545EC	42.2-63.4	52.8	51.2	N/A
LH mU/mL	2545EC	44.2-66.2	55.2	51.8	N/A
	2545EC	44.2-66.2	55.2	52.7	N/A
Progesterone ng/mL	2545EC	23.1-34.7	28.9	29.4	N/A
	2545EC	23.1-34.7	28.9	30.1	N/A

QC TESTED BY: Lisa Lavery
DATE: 3 Oct 25
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *G. McAuley*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016