

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-03-01
KIT DESCRIPTION: IA SPECIALITY I LEV 1
CATALOGUE NO: IAS3113
LOT NUMBERS: 2529EC
EXPIRY DATE: 2027-02-28
SYSTEM OF ANALYSIS: Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Insulin $\mu\text{U/mL}$	2529EC	2.33-3.89	3.11	2.84	N/A
	2529EC	2.33-3.89	3.11	2.77	N/A
	N/A	N/A	N/A	N/A	N/A
Procalcitonin $\mu\text{g/L}$	2529EC	0.749-1.250	0.998	0.986	N/A
	2529EC	0.749-1.250	0.998	0.971	N/A
	N/A	N/A	N/A	N/A	N/A

QC TESTED BY: Lisa Lavery

DATE: 14 Nov 25

RESULT (PASS/FAIL): PASS

QC APPROVED BY:

A. Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016