

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-03-01
KIT DESCRIPTION: IA SPECIALITY I LEV 3
CATALOGUE NO: IAS3115(China)
LOT NUMBERS: 2531EC
EXPIRY DATE: 2027-02-28
SYSTEM OF ANALYSIS: Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Insulin μU/mL	2531EC	22.80-38.00	30.40	29.30	N/A
	2531EC	22.80-38.00	30.40	29.50	N/A
	N/A	N/A	N/A	N/A	N/A
Total Vitamin D nmol/L	2531EC	67.70-113.00	90.30	91.75	N/A
	2531EC	67.70-113.00	90.30	92.25	N/A
	N/A	N/A	N/A	N/A	N/A

QC TESTED BY: Lisa Lavery

DATE: 23 Oct 25

RESULT (PASS/FAIL): PASS

QC APPROVED BY: *Louise McPherson*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016