

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:	2025-01-29
KIT DESCRIPTION:	Immunoassay Specialty II Level 2
CATALOGUE NO:	IAS3118
LOT NUMBERS:	2471EC
EXPIRY DATE:	2027-01-28
SYSTEM OF ANALYSIS:	Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin	2471EC	99.8-166.0	133	144	N/A
pg/ml	2471EC	99.8-166.0	133	139	N/A
Procalcitonin	2471EC	1.55-2.59	2.07	2.22	N/A
μg/l	2471EC	1.55-2.59	2.07	2.14	N/A

QC TESTED BY:	Davina Graham /Lisa Lavery
DATE:	14 Nov 25
RESULT (PASS/FAIL):	Pass
QC APPROVED BY:	a-Manti