

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Radox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-01-29
KIT DESCRIPTION: Immunoassay Specialty II Level 3
CATALOGUE NO: IAS3119
LOT NUMBERS: 2472EC
EXPIRY DATE: 2027-01-28
SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin pg/ml	2472EC	345-575	460	521	N/A
	2472EC	345-575	460	511	N/A
Procalcitonin µg/l	2472EC	11.6-19.4	15.5	16.0	N/A
	2472EC	11.6-19.4	15.5	15.5	N/A

QC TESTED BY: Davina Graham /Lisa Lavery
DATE: 14 Nov 25
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *A. Manton*

Radox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@radox.com I www.radox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Radox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016