

【英文名称】: COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3)

【中文名称】: 凝血质控品 (水平 3)

【货号】 (CAT. NUMBER): CG5023

【批号】 (LOT NUMBER): 431CG

【包装规格】 (SIZE): 12 x 1ml

【到期】 (EXPIRY): 2027-07-28

本产品的准备和使用详见试剂盒内说明书。

Full instructions for the use and preparation of this product are included with the kit.

质控品靶值和范围如下所示。

Target values and ranges are listed below.

ZH: 22 Sep '25 me

Method 所有仪器均值

COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)

批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28

包装规格 (Size): 12 x 1ml

Analyte	分析项目	Unit 单位	Target 靶值	Range		ISD	2SD	Method
				Low 低值	High 高值			
Activated Partial Thromboplastin Time (APTT)	活化部分凝血酶时间 (APTT)	Seconds	58.3	46.6	70.0	5.85	11.7	Beijing Succeeder aPTT
		Seconds	56.0	44.8	67.2	5.60	11.2	Bio-Ksel System APTTs
		Seconds	64.7	51.8	77.6	6.45	12.9	Biolabo Bio-CK APTT
		Seconds	62.3	49.8	74.8	6.25	12.5	Biolabo Bio-Sil (APTT)
		Seconds	56.6	45.3	67.9	5.65	11.3	Biomedica QuikCoag APTT
		Seconds	91.9	73.5	110	9.05	18.1	Cobas aPTT MedS
		Seconds	58.0	46.4	69.6	5.80	11.6	DG Lab APTT
		Seconds	56.4	45.1	67.7	5.65	11.3	Diagon Dia-PTT Liquid
		Seconds	81.0	64.8	97.2	8.10	16.2	Diagon Dia-PTT Lyophilised
		Seconds	68.4	54.7	82.1	6.85	13.7	Dialab, APTT EA Liquid
		Seconds	54.7	43.8	65.6	5.45	10.9	DIRUI aPTT
		Seconds	67.2	53.8	80.6	6.70	13.4	Dutch Diagnostics aPTT
		Seconds	68.2	54.6	81.8	6.80	13.6	Erba Actime APTT
		Seconds	51.5	41.2	61.8	5.15	10.3	Fortress Diagnostics aPTT
		Seconds	73.4	58.7	88.1	7.35	14.7	Helena APTT SI L Minus
		Seconds	73.5	58.8	88.2	7.35	14.7	HemosIL APTT Lyophilised Silica
		Seconds	51.1	40.9	61.3	5.10	10.2	HemosIL APTT-SP liquid
		Seconds	72.7	58.2	87.2	7.25	14.5	HemosIL Synthasil
		Seconds	50.9	40.7	61.1	5.10	10.2	HT-Coag APTT
		Seconds	60.5	48.4	72.6	6.05	12.1	HUMAN Hemostat APTT
Seconds	63.7	51.0	76.4	6.35	12.7	Labitec aPTT		
Seconds	57.2	45.8	68.6	5.70	11.4	Media IVD aPTT		

Method 所有仪器均值

COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)

批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28

包装规格 (Size): 12 x 1ml

Analyte	分析项目	Unit 单位	Target 靶值	Range		ISD	2SD	Method
				Low 低值	High 高值			
Activated Partial Thromboplastin Time (APTT)	活化部分凝血酶时间 (APTT)	Seconds	52.8	42.2	63.4	5.30	10.6	Mindray APTT (Clotting Method)
		Seconds	57.6	46.1	69.1	5.75	11.5	Mindray/Longisland APTT
		Seconds	56.6	45.3	67.9	5.65	11.3	MTI Diagnostics APTT
		Seconds	54.3	43.4	65.2	5.45	10.9	Rayto aPTT Reagent
		Seconds	89.1	71.3	107	8.95	17.9	Roche Cobas APTT Screen
		Seconds	50.0	40.0	60.0	5.00	10.0	Sekisui Coagpia APPT-IN
		Seconds	53.9	43.1	64.7	5.40	10.8	Siemens/Dade Actin
		Seconds	54.2	43.4	65.0	5.40	10.8	Siemens/Dade Actin Cephaloplastin
		Seconds	51.7	41.4	62.0	5.15	10.3	Siemens/Dade Actin FS
		Seconds	46.1	36.9	55.3	4.60	9.20	Siemens/Dade Actin FSL
		Seconds	98.3	78.6	118	9.85	19.7	Siemens/Dade Pathromtin SL
		Seconds	62.8	50.2	75.4	6.30	12.6	Spinreact APTT
		Seconds	64.9	51.9	77.9	6.50	13.0	Stago APTT
		Seconds	66.2	53.0	79.4	6.60	13.2	Stago Cephascreen
		Seconds	65.9	52.7	79.1	6.60	13.2	Stago CK Prest
		Seconds	65.3	52.2	78.4	6.55	13.1	Stago PTT Automate
		Seconds	51.3	41.0	61.6	5.15	10.3	SUNBIO APTT (ELLAGIC ACID)
		Seconds	69.6	55.7	83.5	6.95	13.9	Tcoag TriniCLOT aPTT S
		Seconds	69.2	55.4	83.0	6.90	13.8	Technoclone DAPTTin TC
		Seconds	52.4	41.9	62.9	5.25	10.5	Technoclone Siron APTT
Seconds	68.9	55.1	82.7	6.90	13.8	Teco TEClot APTT		
Seconds	49.1	39.3	58.9	4.90	9.80	Wiener Lab APTT		
Seconds	50.7	40.6	60.8	5.05	10.1	Wiener Lab APTTest ellagic		
Seconds	55.4	44.3	66.5	5.55	11.1	Yumizen G APTT		
Seconds	54.3	43.4	65.2	5.45	10.9	Yumizen G APTT Liq		

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml				Range		范围		
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
Activated Partial Thromboplastin Time (APTT) Ratio	活化部分凝血酶时间 (APTT) 比值	RATIO	1.62	1.30	1.94	0.160	0.320	Bio-Ksel System APTTs
		RATIO	2.20	1.76	2.64	0.220	0.440	Biolabo Bio-CK APTT
		RATIO	2.10	1.68	2.52	0.210	0.420	Erba Actime APTT
		RATIO	1.73	1.38	2.08	0.175	0.350	HemosIL APTT-SP liquid
		RATIO	2.38	1.90	2.86	0.240	0.480	HemosIL Synthasil
		RATIO	2.10	1.68	2.52	0.210	0.420	HUMAN Hemostat APTT
		RATIO	2.09	1.67	2.51	0.210	0.420	Labitec aPTT
		RATIO	1.83	1.46	2.20	0.185	0.370	Mindray APTT (Clotting Method)
		RATIO	1.90	1.52	2.28	0.190	0.380	MTI Diagnostics APTT
		RATIO	3.19	2.55	3.83	0.320	0.640	Roche Cobas APTT
		RATIO	2.03	1.62	2.44	0.205	0.410	Siemens/Dade Actin FS
		RATIO	1.72	1.38	2.06	0.170	0.340	Siemens/Dade Actin FSL
		RATIO	3.14	2.51	3.77	0.315	0.630	Siemens/Dade Pathromtin SL
		RATIO	1.93	1.54	2.32	0.195	0.390	Stago APTT
		RATIO	2.22	1.78	2.66	0.220	0.440	Stago Cephascreen
		RATIO	2.20	1.76	2.64	0.220	0.440	Stago CK Prest
		RATIO	1.91	1.53	2.29	0.190	0.380	Stago PTT Automate
		RATIO	1.83	1.46	2.20	0.185	0.370	SUNBIO APTT (ELLAGIC ACID)
RATIO	2.38	1.90	2.86	0.240	0.480	Tcoag TriniCLOT aPTT S		
RATIO	1.84	1.47	2.21	0.185	0.370	Yumizen G APTT Liq		

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG		货号 (Cat. No): CG5023		到期 (Expiry): 2027-07-28				
包装规格 (Size): 12 x 1ml		Range		范围				
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
Antithrombin III activity	抗凝血酶III活性	%	58.9	47.1	70.7	5.90	11.8	HemosIL Liquid Antithrombin
		%	63.7	51.0	76.4	6.35	12.7	Mindray / Longisland AT-III
		%	61.1	48.9	73.3	6.10	12.2	Roche Antithrombin
		%	62.3	49.8	74.8	6.25	12.5	Roche Chromogenic Thrombin
		%	61.2	49.0	73.4	6.10	12.2	Sclavo ATIII
		%	62.8	50.2	75.4	6.30	12.6	Siemens Innovance Antithrombin III
		%	62.9	50.3	75.5	6.30	12.6	Siemens/Dade Antithrombin III
		%	63.3	50.6	76.0	6.35	12.7	Siemens/Dade Berichrom ATIII
		%	67.0	53.6	80.4	6.70	13.4	STACHrom ATIII
		%	64.0	51.2	76.8	6.40	12.8	SUNBIO AT III
Factor II activity	II因子活性	%	53.9	43.1	64.7	5.40	10.8	HemosIL Factor II deficient plasma
		%	53.4	42.7	64.1	5.35	10.7	HemosIL RecombiPlasTin 2G
		%	61.7	49.4	74.0	6.15	12.3	Siemens Factor II Deficient Plasma
		%	61.2	49.0	73.4	6.10	12.2	Siemens/Dade Innovin
		%	63.0	50.4	75.6	6.30	12.6	Siemens/Dade Thromborel S
		%	56.5	45.2	67.8	5.65	11.3	Stago STA-NeoPTimal
Factor IX activity	IX因子活性	%	56.8	45.4	68.2	5.70	11.4	HemosIL Factor IX deficient plasma
		%	56.0	44.8	67.2	5.60	11.2	HemosIL Synthasil
		%	61.0	48.8	73.2	6.10	12.2	Siemens Factor IX Deficeint Plasma
		%	68.2	54.6	81.8	6.80	13.6	Siemens/Dade Actin FS
		%	57.5	46.0	69.0	5.75	11.5	Siemens/Dade Actin FSL
		%	60.7	48.6	72.8	6.05	12.1	Siemens/Dade Pathromtin SL
		%	55.6	44.5	66.7	5.55	11.1	Stago CK Prest

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG		货号 (Cat. No): CG5023		到期 (Expiry): 2027-07-28				
包装规格 (Size): 12 x 1ml		Range		范围				
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
Factor V activity	V因子活性	%	48.8	39.0	58.6	4.90	9.80	HemosIL Factor V deficient plasma
		%	49.7	39.8	59.6	4.95	9.90	HemosIL RecombiPlasTin 2G
		%	57.2	45.8	68.6	5.70	11.4	Siemens Factor V Deficient Plasma
		%	42.9	34.3	51.5	4.30	8.60	Siemens/Dade Innovin
		%	55.7	44.6	66.8	5.55	11.1	Siemens/Dade Thromborel S
		%	49.9	39.9	59.9	5.00	10.0	Stago Deficient Factor V
		%	46.7	37.4	56.0	4.65	9.30	Stago STA-NeoPTimal
Factor VII activity	VII因子活性	%	76.9	61.5	92.3	7.70	15.4	HemosIL Factor VII deficient plasma
		%	80.1	64.1	96.1	8.00	16.0	HemosIL RecombiPlasTin 2G
		%	75.2	60.2	90.2	7.50	15.0	Siemens Factor VII Deficient Plasma
		%	75.5	60.4	90.6	7.55	15.1	Siemens/Dade Innovin
		%	77.8	62.2	93.4	7.80	15.6	Siemens/Dade Thromborel S
		%	66.4	53.1	79.7	6.65	13.3	Stago Deficient Factor VII
		%	61.3	49.0	73.6	6.15	12.3	Stago STA-NeoPTimal
Factor VIII activity	VIII因子活性	%	27.2	21.8	32.6	2.70	5.40	HEMOSIL FACTOR VIII DEFICIENT PLASMA
		%	26.2	21.0	31.4	2.60	5.20	HemosIL Synthasil
		%	29.1	23.3	34.9	2.90	5.80	Siemens Factor VIII Deficient Plasma
Factor VIII activity	VIII因子活性	%	27.9	22.3	33.5	2.80	5.60	Siemens/Dade Actin FS
		%	28.1	22.5	33.7	2.80	5.60	Siemens/Dade Actin FSL
		%	27.9	22.3	33.5	2.80	5.60	Siemens/Dade Pathromtin SL
		%	25.9	20.7	31.1	2.60	5.20	Stago CK Prest
		%	25.3	20.2	30.4	2.55	5.10	Stago Deficient / ImmunoDef VIII

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG		货号 (Cat. No): CG5023		到期 (Expiry): 2027-07-28				
包装规格 (Size): 12 x 1ml		Range		范围				
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	1SD	2SD	Method
Factor X activity	X因子活性	%	63.6	50.9	76.3	6.35	12.7	HemosIL Factor X deficient plasma
		%	62.4	49.9	74.9	6.25	12.5	HemosIL RecombiPlasTin 2G
		%	66.1	52.9	79.3	6.60	13.2	Siemens Factor X Deficient Plasma
		%	65.3	52.2	78.4	6.55	13.1	Siemens/Dade Innovin
		%	70.5	56.4	84.6	7.05	14.1	Siemens/Dade Thromborel S
		%	53.1	42.5	63.7	5.30	10.6	Stago STA-NeoPTimal
Factor XI activity	XI因子活性	%	35.2	28.2	42.2	3.50	7.00	HemosIL Factor XI deficient plasma
		%	32.4	25.9	38.9	3.25	6.50	Siemens Factor XI Deficient Plasma
		%	25.5	20.4	30.6	2.55	5.10	Siemens/Dade Actin FS
		%	27.8	22.2	33.4	2.80	5.60	Siemens/Dade Actin FSL
		%	35.1	28.1	42.1	3.50	7.00	Stago CK Prest
Factor XII activity	XII因子活性	%	53.8	43.0	64.6	5.40	10.8	HemosIL Factor XII deficient plasma
		%	51.2	41.0	61.4	5.10	10.2	HemosIL Synthasil
		%	62.5	50.0	75.0	6.25	12.5	Siemens Factor XII Deficient Plasma
		%	60.2	48.2	72.2	6.00	12.0	Siemens/Dade Actin FS
		%	60.6	48.5	72.7	6.05	12.1	Siemens/Dade Actin FSL
		%	56.6	45.3	67.9	5.65	11.3	Siemens/Dade Pathromtin SL
		%	50.2	40.2	60.2	5.00	10.0	STA-ImmunoDef XII
		%	50.3	40.2	60.4	5.05	10.1	Stago CK Prest

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml				Range	范围			
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	1SD	2SD	Method
Fibrinogen	纤维蛋白原	g/l	1.27	1.02	1.52	0.125	0.250	SUNBIO Fibrinogen
		mg/dl	127	102	152	12.5	25.0	
		g/l	1.37	1.10	1.64	0.135	0.270	Beijing Succeder Fibrinogen
		mg/dl	137	110	164	13.5	27.0	
		g/l	1.40	1.12	1.68	0.140	0.280	Bio-Ksel Fibrinogen
		mg/dl	140	112	168	14.0	28.0	
		g/l	1.43	1.14	1.72	0.145	0.290	Biolabo Bio-Fibri
		mg/dl	143	114	172	14.5	29.0	
		g/l	1.37	1.10	1.64	0.135	0.270	Cobas Fibrinogen
		mg/dl	137	110	164	13.5	27.0	
		g/l	1.14	0.912	1.37	0.115	0.230	Diagon Dia-FIB
		mg/dl	114	91.2	137	11.5	23.0	
		g/l	1.10	0.880	1.32	0.110	0.220	DIRUI Fibrinogen
		mg/dl	110	88.0	132	11.0	22.0	
		g/l	1.12	0.896	1.34	0.110	0.220	Erba Thrombin Reagent
		mg/dl	112	89.6	134	11.0	22.0	
		g/l	1.35	1.08	1.62	0.135	0.270	Fortress Diagnostics PT with Calcium
		mg/dl	135	108	162	13.5	27.0	
g/l	1.29	1.03	1.55	0.130	0.260	HemosIL Fibrinogen C/C XL		
mg/dl	129	103	155	13.0	26.0			
g/l	1.39	1.11	1.67	0.140	0.280	HemosIL PT-Fibrinogen		
mg/dl	139	111	167	14.0	28.0			
g/l	1.42	1.14	1.70	0.140	0.280	HemosIL PT-Fibrinogen HS Plus		
mg/dl	142	114	170	14.0	28.0			

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG		货号 (Cat. No): CG5023		到期 (Expiry): 2027-07-28				
包装规格 (Size): 12 x 1ml		Range		范围				
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
Fibrinogen	纤维蛋白原	g/l	1.28	1.02	1.54	0.130	0.260	HemosIL QFA (bovine thrombin)
		mg/dl	128	102	154	13.0	26.0	
		g/l	1.02	0.816	1.22	0.100	0.200	HemosIL RecombiPlasTin 2G
		mg/dl	102	81.6	122	10.0	20.0	
		g/l	1.51	1.21	1.81	0.150	0.300	HUMAN Fibrinogen
		mg/dl	151	121	181	15.0	30.0	
		g/l	1.23	0.984	1.48	0.125	0.250	Hyphen Biomed Fibriphen
		mg/dl	123	98.4	148	12.5	25.0	
		g/l	1.37	1.10	1.64	0.135	0.270	Labitec Fibrinogen
		mg/dl	137	110	164	13.5	27.0	
		g/l	1.28	1.02	1.54	0.130	0.260	Media IVD Fibrinogen
		mg/dl	128	102	154	13.0	26.0	
		g/l	1.23	0.984	1.48	0.125	0.250	Mindray Fibrogen (Clotting Method)
		mg/dl	123	98.4	148	12.5	25.0	
		g/l	1.18	0.944	1.42	0.120	0.240	Mindray/Longisland Fibrinogen
		mg/dl	118	94.4	142	12.0	24.0	
		g/l	1.36	1.09	1.63	0.135	0.270	Roche Cobas t411 Fibrinogen
		mg/dl	136	109	163	13.5	27.0	
g/l	1.17	0.936	1.40	0.115	0.230	Roche Cobas t511/t711 Fibrinogen		
mg/dl	117	93.6	140	11.5	23.0			
g/l	1.28	1.02	1.54	0.130	0.260	Scavo Fibrinogen		
mg/dl	128	102	154	13.0	26.0			
g/l	1.22	0.976	1.46	0.120	0.240	Sekisui Coagpia Fibrinogen		
mg/dl	122	97.6	146	12.0	24.0			

Method 所有仪器均值

COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)

批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28

包装规格 (Size): 12 x 1ml

Analyte	分析项目	Unit 单位	Target 靶值	Range		ISD	2SD	Method
				Low 低值	High 高值			
Fibrinogen	纤维蛋白原	g/l	1.20	0.960	1.44	0.120	0.240	Siem/DadeThromboplastin deriv.Fibrinogen
		mg/dl	120	96.0	144	12.0	24.0	
		g/l	0.980	0.784	1.18	0.100	0.200	Siemens/Dade Multifibren U
		mg/dl	98.0	78.4	118	10.0	20.0	
		g/l	1.10	0.880	1.32	0.110	0.220	Siemens/Dade Thrombin
		mg/dl	110	88.0	132	11.0	22.0	
		g/l	1.28	1.02	1.54	0.130	0.260	Spinreact Fibrinogen
		mg/dl	128	102	154	13.0	26.0	
		g/l	1.29	1.03	1.55	0.130	0.260	Stago STA Fibrin-Prest Automate
		mg/dl	129	103	155	13.0	26.0	
		g/l	1.36	1.09	1.63	0.135	0.270	Tcoag TriniCLOT Fibrinogen
		mg/dl	136	109	163	13.5	27.0	
		g/l	1.52	1.22	1.82	0.150	0.300	Technoclone Fibrinogen
		mg/dl	152	122	182	15.0	30.0	
g/l	1.29	1.03	1.55	0.130	0.260	Teco TEClot Fibrinogen		
mg/dl	129	103	155	13.0	26.0			
g/l	1.19	0.952	1.43	0.120	0.240	Yumizen G Fibrinogen		
mg/dl	119	95.2	143	12.0	24.0			
Plasminogen activity	纤溶酶原活性	%	56.2	45.0	67.4	5.60	11.2	HemosIL Plasminogen
		%	60.1	48.1	72.1	6.00	12.0	Siemens/Dade Berichrom Plasminogen
		%	63.6	50.9	76.3	6.35	12.7	Stago Stachrom Plasminogen
Protein C activity	蛋白C活性	%	56	45	67	5.50	11.0	IL HemosIL Protein C
		%	56	45	67	5.50	11.0	Siemens/Dade Berichrom Protein C
		%	65	52	78	6.50	13.0	Siemens/Dade Protein C Clotting
		%	61	49	73	6.00	12.0	STACHrom Protein C

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml				Range	范围			
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
Protein S activity	蛋白S活性	%	71	50	92	10.5	21.0	IL HemosIL ProS
		%	58	41	75	8.50	17.0	IL HemosIL Protein S Activity
		%	60	42	78	9.00	18.0	Siemens/Dade Protein S Ac
		%	55	39	71	8.00	16.0	STA Liatest
		%	69	48	90	10.5	21.0	Stago StaClot
Prothrombin Time (PT)	凝血酶原时间 (PT)	Seconds	19.0	15.2	22.8	1.90	3.80	SUNBIO PT
		Seconds	21.0	16.8	25.2	2.10	4.20	Beijing Succeder PT
		Seconds	16.3	13.0	19.6	1.65	3.30	Bio-Ksel PT Plus
		Seconds	15.8	12.6	19.0	1.60	3.20	Bio-Ksel System PT
		Seconds	28.6	22.9	34.3	2.85	5.70	Biolabo Bio-TP
		Seconds	27.9	22.3	33.5	2.80	5.60	Biolabo Bio-TP (Low ISI)
Prothrombin Time (PT)	凝血酶原时间 (PT)	Seconds	14.7	11.8	17.6	1.45	2.90	DG Lab PT
		Seconds	13.8	11.0	16.6	1.40	2.80	Dirui PT
		Seconds	22.3	17.8	26.8	2.25	4.50	Erba Protime PT
		Seconds	24.0	19.2	28.8	2.40	4.80	Helena Thromboplastin LI
		Seconds	16.1	12.9	19.3	1.60	3.20	HemosIL PT-Fibrinogen
		Seconds	26.5	21.2	31.8	2.65	5.30	HemosIL PT-Fibrinogen HS Plus
		Seconds	17.6	14.1	21.1	1.75	3.50	HemosIL REDIPLAS Tin
		Seconds	16.3	13.0	19.6	1.65	3.30	HemosIL RecombiPlasTin 2G
		Seconds	24.3	19.4	29.2	2.45	4.90	HUMAN Thromboplastin
		Seconds	24.7	19.8	29.6	2.45	4.90	Labitec PT
		Seconds	26.0	20.8	31.2	2.60	5.20	Media IVD PT
		Seconds	20.8	16.6	25.0	2.10	4.20	Mindray PT (Clotting Method)
		Seconds	21.4	17.1	25.7	2.15	4.30	Mindray/Longisland PT

Method 所有仪器均值

COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)

批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28

包装规格 (Size): 12 x 1ml

Analyte	分析项目	Unit 单位	Target 靶值	Range		ISD	2SD	Method
				Low 低值	High 高值			
Prothrombin Time (PT)	凝血酶原时间 (PT)	Seconds	24.2	19.4	29.0	2.40	4.80	Rayto Prothrombin Time Reagent
		Seconds	40.3	32.2	48.4	4.05	8.10	Roche Cobas PT Rec
		Seconds	17.8	14.2	21.4	1.80	3.60	Roche Cobas PT Screen
		Seconds	23.9	19.1	28.7	2.40	4.80	Sclavo PT
		Seconds	36.8	29.4	44.2	3.70	7.40	Sclavo PT-HRTF
		Seconds	27.5	22.0	33.0	2.75	5.50	Sekisui Coagpia PT-N
		Seconds	29.2	23.4	35.0	2.90	5.80	Siemens/Dade Innovin
		Seconds	17.2	13.8	20.6	1.70	3.40	Siemens/Dade Thromborel IS
		Seconds	16.8	13.4	20.2	1.70	3.40	Siemens/Dade Thromborel S
		Seconds	29.2	23.4	35.0	2.90	5.80	Stago Neoplastine CI
		Seconds	30.9	24.7	37.1	3.10	6.20	Stago Neoplastine CL Plus -Unicalibrator
		Seconds	19.7	15.8	23.6	1.95	3.90	Stago Neoplastine R
		Seconds	25.1	20.1	30.1	2.50	5.00	Stago STA Neoplastine CI Plus
		Seconds	30.5	24.4	36.6	3.05	6.10	Stago STA-NeoPTimal
		Seconds	45.1	36.1	54.1	4.50	9.00	Tcoag TriniCLOT PT Excel S
		Seconds	27.9	22.3	33.5	2.80	5.60	Technoclone Technoplastin HIS
		Seconds	18.9	15.1	22.7	1.90	3.80	Technology-Standard PT
		Seconds	22.3	17.8	26.8	2.25	4.50	Teco TEClot PT-S
		Seconds	33.8	27.0	40.6	3.40	6.80	Wiener Lab Soluplastin
		Seconds	23.0	18.4	27.6	2.30	4.60	Yumizen G PT
Seconds	33.5	26.8	40.2	3.35	6.70	Yumizen G PT Liq		

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG		货号 (Cat. No): CG5023		到期 (Expiry): 2027-07-28				
包装规格 (Size): 12 x 1ml				Range		范围		
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	1SD	2SD	Method
Prothrombin Time (PT) activity	凝血酶原时间 (PT) 活性	%	39.8	31.8	47.8	4.00	8.00	SUNBIO PT
		%	41.8	33.4	50.2	4.20	8.40	Beijing Succeder PT
		%	35.7	28.6	42.8	3.55	7.10	Biolabo Bio-TP
		%	39.0	31.2	46.8	3.90	7.80	Biolabo Bio-TP (Low ISI)
		%	34.4	27.5	41.3	3.45	6.90	Diagon, Dia-PT
		%	37.4	29.9	44.9	3.75	7.50	HemosIL PT-Fibrinogen HS Plus
		%	55.4	44.3	66.5	5.55	11.1	HemosIL ReadiPlasTin
		%	58.5	46.8	70.2	5.85	11.7	HemosIL RecombiPlasTin 2G
		%	36.6	29.3	43.9	3.65	7.30	Labitec PT
Prothrombin Time (PT) activity	凝血酶原时间 (PT) 活性	%	40.2	32.2	48.2	4.00	8.00	Media IVD PT
		%	13.2	10.6	15.8	1.30	2.60	Roche Cobas PT Rec
		%	44.6	35.7	53.5	4.45	8.90	Roche Cobas PT Screen
		%	40.6	32.5	48.7	4.05	8.10	Scavo PT
		%	19.9	15.9	23.9	2.00	4.00	Siemens/Dade Innovin
		%	54.0	43.2	64.8	5.40	10.8	Siemens/Dade Thromborel S
		%	32.6	26.1	39.1	3.25	6.50	Stago Neoplastine CI
		%	53.4	42.7	64.1	5.35	10.7	Stago Neoplastine R
		%	39.7	31.8	47.6	3.95	7.90	Stago STA Neoplastine CI Plus
		%	31.8	25.4	38.2	3.20	6.40	Stago STA-NeoPTimal
		%	17.9	14.3	21.5	1.80	3.60	Tcoag TriniCLOT PT Excel S
		%	36.6	29.3	43.9	3.65	7.30	Teco TEClot PT-S
		%	19.3	15.4	23.2	1.95	3.90	Yumizen G PT Liq

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml				Range	范围			
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
PT as a ratio	凝血酶原时间 (PT) 比值	RATIO	1.58	1.26	1.90	0.160	0.320	SUNBIO PT
		RATIO	1.44	1.15	1.73	0.145	0.290	HemosIL ReadiPlasTin
		RATIO	1.43	1.14	1.72	0.145	0.290	HemosIL RecombiPlasTin 2G
		RATIO	1.76	1.41	2.11	0.175	0.350	HUMAN Thromboplastin
		RATIO	1.88	1.50	2.26	0.190	0.380	Sclavo PT
		RATIO	2.84	2.27	3.41	0.285	0.570	Siemens/Dade Innovin
		RATIO	1.44	1.15	1.73	0.145	0.290	Siemens/Dade Thromborel S
		RATIO	1.69	1.35	2.03	0.170	0.340	Stago STA Neoplastine CI Plus
		RATIO	2.27	1.82	2.72	0.225	0.450	Stago STA-NeoPTimal
PT as an INR	凝血酶原时间 (PT) 国际标准化比值	INR	1.61	1.29	1.93	0.160	0.320	SUNBIO PT
		INR	1.71	1.37	2.05	0.170	0.340	Beijing Succeeder PT
		INR	1.49	1.19	1.79	0.150	0.300	Bio-Ksel PT Plus
		INR	2.29	1.83	2.75	0.230	0.460	Biolabo Bio-TP
		INR	2.23	1.78	2.68	0.225	0.450	Biolabo Bio-TP (Low ISI)
		INR	1.92	1.54	2.30	0.190	0.380	Biomedica QuikCoag PT-HS
		INR	1.24	0.992	1.49	0.125	0.250	DG Lab PT
		INR	1.98	1.58	2.38	0.200	0.400	Diagon, Dia-PT
		INR	1.23	0.984	1.48	0.125	0.250	Dirui PT
		INR	1.86	1.49	2.23	0.185	0.370	Helena Thromboplastin LI
		INR	1.41	1.13	1.69	0.140	0.280	HemosIL PT-Fibrinogen
		INR	2.00	1.60	2.40	0.200	0.400	HemosIL PT-Fibrinogen HS Plus
		INR	1.57	1.26	1.88	0.155	0.310	HemosIL ReadiPlasTin
		INR	1.45	1.16	1.74	0.145	0.290	HemosIL RecombiPlasTin 2G
		INR	2.19	1.75	2.63	0.220	0.440	HUMAN Thromboplastin
		INR	6.16	4.93	7.39	0.615	1.23	Hyphen Biomed Thrombophen
INR	2.15	1.72	2.58	0.215	0.430	Labitec PT		

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml		Range		范围				
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	Method
PT as an INR	凝血酶原时间 (PT) 国际标准化比值	INR	2.07	1.66	2.48	0.205	0.410	Media IVD PT
		INR	1.76	1.41	2.11	0.175	0.350	Mindray/Longisland PT
		INR	4.13	3.30	4.96	0.415	0.830	Roche Cobas PT Rec
		INR	1.85	1.48	2.22	0.185	0.370	Roche Cobas PT Screen
		INR	2.08	1.66	2.50	0.210	0.420	Sclavo PT
		INR	3.45	2.76	4.14	0.345	0.690	Sclavo PT-HRTF
		INR	2.28	1.82	2.74	0.230	0.460	Sekisui Coagpia PT-N
		INR	2.87	2.30	3.44	0.285	0.570	Siemens/Dade Innovin
		INR	1.35	1.08	1.62	0.135	0.270	Siemens/Dade Thromborel IS
		INR	1.47	1.18	1.76	0.145	0.290	Siemens/Dade Thromborel S
		INR	2.54	2.03	3.05	0.255	0.510	Spinreact PT
		INR	2.32	1.86	2.78	0.230	0.460	Stago Neoplastine CI
		INR	1.44	1.15	1.73	0.145	0.290	Stago Neoplastine R
		INR	2.08	1.66	2.50	0.210	0.420	Stago STA Neoplastine CI Plus
		INR	2.44	1.95	2.93	0.245	0.490	Stago STA-NeoPTimal
		INR	4.57	3.66	5.48	0.455	0.910	Tcoag TriniCLOT PT Excel S
		INR	1.71	1.37	2.05	0.170	0.340	Technoclone Technoplastin HIS
		INR	1.50	1.20	1.80	0.150	0.300	Technology-Standard PT
		INR	1.75	1.40	2.10	0.175	0.350	Teco TEClot PT-S
		INR	3.12	2.50	3.74	0.310	0.620	Wiener Lab Soluplastin
INR	1.66	1.33	1.99	0.165	0.330	Yumizen G INR kit		
INR	2.04	1.63	2.45	0.205	0.410	Yumizen G PT Reco		

Method 所有仪器均值		COAGULATION CONTROL - LEVEL 3 (COAG CONTROL 3) 凝血质控品 (水平3)						
批号 (Lot. No): 431CG 货号 (Cat. No): CG5023 到期 (Expiry): 2027-07-28								
包装规格 (Size): 12 x 1ml				Range		范围		
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	1SD	2SD	Method
Thrombin Time (TT)	凝血酶时间 (TT)	Seconds	26.5	21.2	31.8	2.65	5.30	SUNBIO TT
		Seconds	40.1	32.1	48.1	4.00	8.00	Beijing Succeder TT
		Seconds	34.6	27.7	41.5	3.45	6.90	Biolabo Bio-TT
		Seconds	32.7	26.2	39.2	3.25	6.50	Erba Thrombin Time
		Seconds	31.3	25.0	37.6	3.15	6.30	HemosIL Thrombin Time 5ml
		Seconds	37.3	29.8	44.8	3.75	7.50	HemosIL Thrombin Time 8ml
		Seconds	27.6	22.1	33.1	2.75	5.50	Mindray TT (Clotting Method)
		Seconds	24.2	19.4	29.0	2.40	4.80	Mindray/Longisland TT
Thrombin Time (TT)	凝血酶时间 (TT)	Seconds	39.7	31.8	47.6	3.95	7.90	Siemens BC Thrombin
		Seconds	33.2	26.6	39.8	3.30	6.60	Siemens Test Thrombin
		Seconds	32.9	26.3	39.5	3.30	6.60	Siemens/Dade Thrombin
		Seconds	30.9	24.7	37.1	3.10	6.20	Siemens/Dade Thromboclotin
		Seconds	39.4	31.5	47.3	3.95	7.90	Stago Thrombin
		Seconds	32.3	25.8	38.8	3.25	6.50	Sysmex Thrombin Time