

【英文名称】: CSF CONTROL - LEVEL 2 (CSF CONTROL 2)

【中文名称】: 脑脊液质控品（水平 2）

【货号】 (CAT. NUMBER): CFI500

【批号】 (LOT NUMBER): 231CF

【包装规格】 (SIZE): 10 x 3ml

【到期】 (EXPIRY): 2026-11-28

本产品的准备和使用详见试剂盒内说明书。

Full instructions for the use and preparation of this product are included with the kit.

质控品靶值和范围如下所示。

Target values and ranges are listed below.

ZH: Rev. 25 Jan '24 me

Method 所有仪器均值		CSF CONTROL - LEVEL 2 (CSF CONTROL 2)					脑脊液质控品 (水平2)			
批号 (Lot. No): 231CF		货号 (Cat. No): CF1500		到期 (Expiry): 2026-11-28						
包装规格 (Size): 10 x 3ml		Range		范围						
Analyte	分析项目	Unit 单位	Target 靶值	Low 低值	High 高值	ISD	2SD	方法	Method	
Chloride	氯	mmol/l	90.5	83.3	97.7	3.60	7.20	离子选择性电极法	Ion Selective Electrode	
Glucose	葡萄糖	mmol/l	3.27	2.78	3.76	0.245	0.490	葡糖氧化酶法	Glucose oxidase	
		mg/dl	58.9	50.1	67.7	4.40	8.80			
		mmol/l	3.30	2.81	3.79	0.245	0.490	己糖激酶法	Hexokinase	
		mg/dl	59.5	50.6	68.4	4.45	8.90			
IgG	免疫球蛋白G (IgG)	mg/dl	4.00	3.28	4.72	0.360	0.720	免疫比浊法	Immunoturbidimetric	
		mg/l	40.0	32.8	47.2	3.60	7.20			
Lactate	乳酸	mmol/l	0.441	0.362	0.520	0.040	0.079	酶法	Enzymatic	
		mg/dl	3.97	3.26	4.68	0.355	0.710			
Protein, Total	总蛋白	g/l	0.278	0.222	0.334	0.028	0.056	邻苯三酚红	Pyrogallol Red	
		mg/dl	27.8	22.2	33.4	2.80	5.60			
		mg/l	278	222	334	28.0	56.0			
Sodium	钠	mmol/l	130	124	136	3.00	6.00	离子选择性电极法	Ion Selective Electrode	