



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2024-07-29
Product Description:	COAGULATION CONTROL LEVEL 3
Catalogue Number:	CG5023
Batch Number:	431CG
Expiry Date:	2027-07-28

Analyte	Lot No.	Target	Range		Result	Units
APTT	431CG	72.7	58.2	87.2	67.5	Seconds

Approved By:

L. Pandharani

Date: 12 Nov '25

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
+44 (0) 28 9442 2413 | marketing@randox.com | randox.com

Directors: Dr S P FitzGerald CBE FREng DL BSc PhD DSc (Hons) FIAE Man. R P Kelly BSc FCA.

Randox Laboratories Ltd is a subsidiary of Randox Holdings Limited a company registered within Northern Ireland with company number NI. 614690. VAT Registered Number: GB 151 6827 08.

CG5023 Ch
Revision 0

7677-QC
Revision (4)
Last Printed Date: 12/11/2025