

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-08-29  
KIT DESCRIPTION: Immunoassay Premium Plus Level 1 (IA PREMIUM PLUS1)  
CATALOGUE NO: IA3109  
LOT NUMBERS: 2651EC  
EXPIRY DATE: 2028-08-28  
SYSTEM OF ANALYSIS: Roche e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
TT3 nmol/L	2651EC	2.02-3.36	2.69	3.02	N/A
	2651EC	2.02-3.36	2.69	3.07	N/A
LH mU/mL	2651EC	2.98-4.46	3.72	4.09	N/A
	2651EC	2.98-4.46	3.72	3.98	N/A
Progesterone ng/mL	2651EC	0.771-1.160	0.964	0.944	N/A
	2651EC	0.771-1.160	0.964	0.974	N/A

QC TESTED BY: Lisa Lavery  
DATE: 24 Oct 25  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: *A. McArthur*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016