

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-08-29
KIT DESCRIPTION: Immunoassay Premium Plus Level 3 (IA PREMIUM PLUS 3)
CATALOGUE NO: IA3111
LOT NUMBERS: 2681EC
EXPIRY DATE: 2028-08-28
SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH mU/mL	2681EC	44.4-66.6	55.5	53.6	N/A
	2681EC	44.4-66.6	55.5	52.8	N/A
LH mU/mL	2681EC	41.8-62.8	52.3	51.9	N/A
	2681EC	41.8-62.8	52.3	51.4	N/A
Progesterone ng/mL	2681EC	22.8-34.2	28.5	27.8	N/A
	2581EC	22.8-34.2	28.5	28.1	N/A

QC TESTED BY: LISA LAVERY
DATE: 12 Dec 25
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *A. Menden*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016