

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-07-29
KIT DESCRIPTION: Immunoassay Premium Plus Level 2(IA PREMIUM PLUS 2) China
CATALOGUE NO: IA3110
LOT NUMBERS: 2544EC
EXPIRY DATE: 2028-07-28
SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH mU/mL	2544EC	19.4-29.0	24.2	24.3	N/A
	2544EC	19.4-29.0	24.2	24.0	N/A
LH mU/mL	2544EC	30.5-45.7	38.1	35.8	N/A
	2544EC	30.5-45.7	38.1	36.0	N/A
Progesterone ng/mL	2544EC	6.63-9.95	8.29	8.44	N/A
	2544EC	6.63-9.95	8.29	8.79	N/A

QC TESTED BY: Lisa Lavery
DATE: 3 Oct 25
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *A. Montin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016