

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-01-29
KIT DESCRIPTION: Immunoassay Specialty II Level 1
CATALOGUE NO: IAS3117
LOT NUMBERS: 2470EC
EXPIRY DATE: 2027-01-28
SYSTEM OF ANALYSIS: Roche e801/Immulinite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin pg/ml	2470EC	21.1-35.1	28.1	29.1	N/A
	2470EC	21.1-35.1	28.1	29.6	N/A
Gastrin pmol/l	2470EC	8.85-14.8	11.8	12.1	N/A
	2470EC	8.85-14.8	11.8	12.8	N/A

QC TESTED BY: Davina Graham /Maya Pulikunni Vijayan
DATE: 29-Jan-26
RESULT (PASS/FAIL): Pass
QC APPROVED BY: A. Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016