RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-08-29

KIT DESCRIPTION: Immunoassay Premium Plus Level 3(IA PREMIUM PLUS 3)

CATALOGUE NO: IA3111

LOT NUMBERS: 2653EC

EXPIRY DATE: 2028-08-28

SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH mU/mL	2653EC	44.4-66.6	55.5	55.0	N/A
	2653EC	44.4-66.6	55.5	54.5	N/A
LH mU/mL	2653EC	41.8-62.8	52.3	55.2	N/A
	2653EC	41.8-62.8	52.3	54.5	N/A
Progesterone ng/mL	2653EC	22.8-34.2	28.5	28.7	N/A
	2653EC	22.8-34.2	28.5	27.6	N/A

QC TESTED BY: MAYA PULIKUNNI VIJAYAN

DATE: 06-Nov-25

RESULT (PASS/FAIL): Pass

QC APPROVED BY: A. Mantes

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com