BLOOD GAS CONTROL - LEVEL 3 (BG CONTROL 3)

CAT. NO. BG5003
SIZE: 30 x 1.8 ml
GTIN: 05055273227123
LOT NO. 321BG
EXPIRY: 2027-11-28

INTENDED USE
This product is intended for in vitro diagnostic use, in the quality control of Blood Gas analysis.

DEVICE DESCRIPTION
The Blood Gas Controls are supplied at 3 levels, 1, 2 and 3. Target values and ranges are supplied for the following analytes: Calcium, Chloride, Glucose, Lactate, PCO_2, pH, pO_2, Potassium, Sodium and Total CO_2.

SAFETY PRECAUTIONS AND WARNINGS
For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents. Health and Safety Data Sheets are available on request.

Not suitable for instruments that do not recommend products with artificial dye.

STORAGE AND STABILITY
UNOPENED: The product is stable to expiration date when stored at +2°C to +8°C. Avoid exposure to freezing and temperatures greater than +30°C.

OPENED: For pH/blood gas values, the control should be analysed within 1 minute of opening. For electrolyte measurements, the control should be analysed within 1 hour after opening.

PREPARATION FOR USE
The Blood Gas Control should be brought to +20°C to +23°C before use. Allow at least 4 hours for ampoules to equilibrate to this temperature, prior to testing. Before use, hold the ampoule at the top and bottom (with forefinger and thumb) and shake 15 - 20 times to mix the solution. Tap the ampoule to restore the liquid to the bottom of the ampoule. Open the ampoule by snapping off the tip at the score. Use gauze, tissue, gloves or an appropriate ampoule opener to protect fingers from cuts. Immediately introduce the liquid from the ampoule to the analyser.

MATERIALS PROVIDED
Blood Gas Control - Level 3 30 x 1.8 ml

ASIGNED VALUES
Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

Each batch of Blood Gas Control is submitted to a number of external laboratories and values are assigned from a consensus of results obtained by these laboratories.

EC REP Randox Teoranta, Meenmore,
Dungloe, Donegal,
F94 TV06, Ireland

08 Apr '26 bm

Table of Content

Method 2

Randox Laboratories Ltd., 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
Tel: +44 (0) 28 9442 2413 Fax: +44 (0) 28 9445 2912
Email: applications@randox.com Website: www.randox.com

All enzyme values are quoted to a temperature of 37°C, unless otherwise stated

RANDOX

Method

Blood Gas Control - Level 3 (BG CONTROL 3)

Lot. No: 321BG Cat. No: BG5003 Expiry: 2027/11/28

Size: 30 x 1.8 ml

			Range				
Analyte	**Unit**	**Target**	**Low**	**High**	**1SD**	**2SD**	**Method**
Calcium	mg/dl	2.61	2.35	2.87	0.130	0.260	Ion Selective Electrode
	mmol/l	0.650	0.585	0.715	0.033	0.065	
Chloride	mmol/l	124	114	134	5.00	10.0	ISE Indirect
Glucose	mg/dl	252	214	290	19.0	38.0	Enzymatic Electrode
	mmol/l	14.0	11.9	16.1	1.05	2.10	
pCO2	kPa	2.61	2.09	3.13	0.260	0.520	Ion Selective Electrode
pH	pH units	7.60	7.52	7.68	0.040	0.080	Ion Selective Electrode
pO2	kPa	19.3	15.4	23.2	1.95	3.90	Ion Selective Electrode
Potassium	mmol/l	5.99	5.51	6.47	0.240	0.480	Ion Selective Electrode Method
Sodium	mmol/l	158	150	166	4.00	8.00	Ion Selective Electrode Method
Total CO2	mmol/l	18.7	15.0	22.4	1.85	3.70	Ion Selective Electrode

Randox Laboratories Ltd., 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom

Tel: +44 (0) 28 9442 2413 Fax: +44 (0) 28 9445 2912

Email: applications@randox.com Website: www.randox.com

All enzyme values are quoted to a temperature of 37°C, unless otherwise stated