RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2026-05-29

KIT DESCRIPTION: Immunoassay Premium Plus Level 1(IA PREMIUM PLUS 1)

CATALOGUE NO: IA3109

LOT NUMBERS: 2719EC

EXPIRY DATE: 2029-05-28

SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
LH mU/mL	2719EC	2.82 – 4.22	3.52	3.35	N/A
	2719EC	2.82 – 4.22	3.52	3.27	N/A
Prolactin µIU/mL	2719EC	110 - 166	138	134	N/A
	2719EC	110 - 166	138	132	N/A
Progesterone ng/mL	2719EC	0.776 – 1.16	0.970	1.030	N/A
	2719EC	0.776 – 1.16	0.970	1.120	N/A

QC TESTED BY: Maya Pulikunni Vijayan

DATE: 18 Jun 26

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.