RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2025-11-29

KIT DESCRIPTION: Immunoassay Premium Plus - Level 2 (IA PREMIUM PLUS 2)

CATALOGUE NO: IA3110

LOT NUMBERS: 2659EC/2

EXPIRY DATE: 2028-11-28

SYSTEM OF ANALYSIS: Roche e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
LH mIU/mL	2659EC/2	26.2-39.2	32.7	30.5	N/A
	2659EC/2	26.2-39.2	32.7	30.9	N/A
Prolactin µIU/mL	2659EC/2	559-839	699	666	N/A
	2659EC/2	559-839	699	675	N/A
Progesterone ng/mL	2659EC/2	6.83-10.30	8.54	9.57	N/A
	2659EC/2	6.83-10.30	8.54	9.08	N/A

QC TESTED BY: Maya Pulikunni Vijayan

DATE: 30-Jun-26

RESULT (PASS/FAIL): Pass

QC APPROVED BY: R. Marten

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc FhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.