RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	**2025-09-29**
KIT DESCRIPTION:	**Lipid Control Level III**
CATALOGUE NO:	**LE2663**
BATCH NO:	**3222CH**
EXPIRY DATE:	**2028-09-28**

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 3222CH	HDL-Cholesterol	1.59	1.35-1.83	1.57	mmol/l
Level III / 3222CH	LDL - Cholesterol	3.98	3.38-4.58	4.15	mmol/l

APPROVED BY: A. Rycroft

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM No. 9323-QC
REVISION (1)
27th September 2013